

TO: Kristie Warr

FROM: Rick Haaker, CHP, CIH

SUBJECT: Review of Las Conchas Fire Work Order 11-07013

DATE: 7/22/2011

Discussion concerning U-235 results is provided as the last paragraph in this revision.

The data were reviewed for accuracy, completeness, and any apparent issues. During data review a qualifier "UB" was assigned if the activity result is less than five times the activity result of the method blank. A "UB" qualifier denotes that an analyte is non-detect due to lack of activity relative to a blank concentration. Unused filters from the same lot as the sample filters were used as the method blank. The analytes in Table 1 were detected in the method blank, and all samples are affected.

Table 1. Analytes detected or tentatively detected in the method blank and data qualifiers based on the blank.

BlanksWithDetectedOrEstimatedActivity	
Isotope	Assigned Qualifier
GROSS ALPHA	J
GROSS BETA	
PU-238	J
PU-239	J
TOTAL SR	
U-234	
U-235	J
U-238	

Data without a UB qualifier was further reviewed.

A "U" was assigned to the Assigned Qualifier column when result was less than 50% of the MDA. In this case the analytical result was assigned to be one-half of the MDA in the "ValidatedResult" column. The validated result should be considered an upper bound estimate in this case.

A "J" was assigned if the result was between 0.5 of the MDA and the MDA. The validated result is the reported result. The validated result represents an

estimated value in this case.

A “JH” or “JL” would be based on percent recovery (the "RadioPercentRec", “GravPercentRec” and the “LCSpercentR” columns of the Eberline Services report. Recoveries below 70% would result in a JL to denote the result is estimated with a low bias. Recoveries above 130% would result in assignment of a JH to denote that the result is estimated with a high bias.

Table 2 lists samples that exhibited percent recoveries outside of the acceptance range. These samples all carried a “U” or “UB” qualifier.

Table 2. Samples with recoveries outside of the acceptance range.

Isotope	ClientID	RadioPercentRec
AM-241	A005-110630-1801-1-T01	52.43
PU-238	A001-110701-1505-1-T01	68.96
PU-238	A003-110630-ST01	53.16
PU-238	A006-110630-1702-1-T01	61.13
PU-239	A001-110701-1505-1-T01	68.96
PU-239	A003-110630-ST01	53.16
PU-239	A006-110630-1702-1-T01	61.13
U-234	A002-110630-1732-1-T01	135.54
U-234	A001-110701-1505-1-T01	133.52
U-234	A003-110630-1632-1-T01	136.87
U-234	A006-110630-ST01	141.24
U-235	A002-110630-1732-1-T01	135.54
U-235	A001-110701-1505-1-T01	133.52
U-235	A003-110630-1632-1-T01	136.87
U-235	A006-110630-ST01	141.24
U-238	A002-110630-1732-1-T01	135.54
U-238	A001-110701-1505-1-T01	133.52
U-238	A003-110630-1632-1-T01	136.87
U-238	A006-110630-ST01	141.24

The assigned data qualifiers are found in column “AssignedQualifier”.

The effective air volume for the various analytes of the various air samples in cubic meters are provided in the column “AliquotNetEquiv”.

Note that the blank results are in pCi/m³. The volume that Eberline Services assigned to the blanks for a given analyte are the average of the effective volumes for the samples in the sample set for that analyte.

Air volumes that were collected in this sample set tend to be lower than ideal for environmental samples, being in the range of 36 to 101 cubic meters.

The period of time between collection of air samples and gross alpha/beta counting was short, ranging from one to three days, so those results are likely to include an activity contribution due to the presence of radon daughters.

The EDD provided by Eberline Services contains erroneous sample dates that do not agree with the collection dates provided on the chain of custody No: 6-070111-162204-002. No discrepancies were found in the transcription of sample IDs or sample volumes from the chain of custody to the EDD.

Samples A004-110630-ST01 and A005-110630-ST01 did not require a UB qualifier for U-235 because the activity exceeded 5 times the measured value in the blank. Both of these results carry a high uncertainty and the results do not have the blank activity subtracted from them. If the U-235 activity of the blank were subtracted from that of the sample and the associated error estimates are propagated through the calculation, then the net U-235 activity is not likely differ significantly from zero.